
HPLC Application

ID No.: **14751**

Anion-Exchange Std. (AL0-3041)



Column: PhenoSphere 5u SAX 80A

Dimensions: 250 x 4.6 mm ID

Order No: 00G-3151-E0

Elution Type: Isocratic


Eluent A: 0.15M (NH4)2 HPO4, pH 6.0 **Pct. A:** 100

Flow Rate: 0.6 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs. - Variable Wave. @ 254 nm (ambient)

Analyst Note: This standard is used to test the following Phenomenex silica-based ion-exchange columns:
SAX, SB, DEAE, PEI.
NOTE: Separation conditions and results may vary for your column. Contact your Technical Representative.



ANALYTES:

1 Uridine

2 UMP

©Copyright 2000-2007 Phenomenex Inc.